

Stock Code: 300298



iCARE-2100

Portable Automatic Multi-function Analyzer

Core Technology
— **Liquid Phase iPOCT**



■ ABOUT SINOCARE

By cherishing our core values of “Honor the Commitment, Dedicate to Healthcare”, we have been fully dedicated to the innovation of biosensor technology, developing, manufacturing and marketing rapid diagnosis testing products for people with chronic diseases and healthcare professionals.

Sinocare

iCARE-2100

Portable Automatic Multi-function Analyzer




iCARE-2100



Cartridge holder

- Accuracy: comparable result
- Multi-parameters: including biochemistry, coagulation, HbA1c, ACR parameters
- Efficiency: testing time within 3-15 minutes, iPOCT system is exceedingly suit for individual test, truly on demand
- Easy of operation: only three steps, no need of professional operation and calibration
- Economy: low cost without additional consumables & no daily maintenance required
- Compatibility: compatible with your data management system
- Self-contained and single-use cartridge, unique tips

21* cartridges 51* indicators



Linearity Within Optical Measurement Range

The linearity within optical measurement range meets the requirement below:

Measuring system	Linearity correlation coefficient
Light transmission system	The maximum light absorbency =2.0,R=0.995
Fluorescence system	Fluorescent intensity within 500-2300,R=0.99
Light reflection system	Reflectivity within 5.5%~80.0%,R=0.99

Sinocare Rapid Reagent Kit

Reagent Kit	Detection Indicator	Detection Indicator	Test Principle	Specimen	Volume	Time			
Inflammation I	FR CRP(hs-CRP+CRP)	C-Reactive Protein	Latex immunoturbidimetry	Whole blood	15 µl	6min			
Inflammation II	FR CRP(hs-CRP+CRP)	C-Reactive Protein	Latex immunoturbidimetry	Serum	150 µl	9min			
	SAA	Serum Amyloid A							
Liver Function I	ALT	Alanine Aminotransferase	Rate method	Serum	150 µl	15min			
	AST	Aspartate Aminotransferase							
	T-Bil	Total Bilirubin	End-point method						
	ALB	Albumin							
Liver Function II	D-Bil	Direct Bilirubin	End-point method	Serum	150 µl	14min			
	TP	Total Protein							
	ALP	Alkaline Phosphatase	Rate method						
	GGT	γ-Glutamyl Transferase							
Blood lipid	TC	Total cholesterol	Enzymatic method (End-point method)	Serum	150 µl	15min			
	TG	Triglycerides							
	HDL-C	HDL cholesterol							
	LDL-C	LDL cholesterol							
Renal Function I	Urea	Urea	Enzymatic kinetic method	Serum	150 µl	12min			
	Crea	Creatinine	Enzymatic method (End-point method)						
	UA	Uric Acid							
Renal Function II	β 2-MG	β 2-Microglobulin	Latex immunoturbidimetry	Serum	150 µl	13min			
	Cys-C	Cystatin C							
Screening for chronic disease	Glu	Glucose	Enzymatic method	Serum	150 µl	13min			
	HCY	Homocysteine							
	UA	Uric Acid							
	LDL-C	Low Density Lipoprotein Cholesterol							
Coagulation	PT	Prothrombin time	Clotting method	Plasma	400 µl	15min			
	APTT	Activated Partial Thromboplastin time							
	TT	Thrombin time							
	FIB	Fibrinogen							

Reagent Kit	Detection Indicator	Detection Indicator	Test Principle	Specimen	Volume	Time
Prothrombin time /INR	PT	Prothrombin time	Clotting method	Plasma	150 µl	5min
	INR	INR	/			
D-Dimer	D-Dimer	D-Dimer	Latex immunoturbidimetry	Plasma	150 µl	9min
Myocardial Enzyme	CK	Creatine kinase	Rate method	Serum	150 µl	14min
	CK-MB	Creatine kinase myocardial band				
	LDH	Lactate dehydrogenase				
	α HBDH	α - Hydroxybutyrate Dehydrogenase				
Mellitus I	Glu	Glucose	Enzymatic method	Serum	150 µl	9min
	GA	Glycated Serum Albumin				
HbA1c	HbA1c	Glycated Hemoglobin	Boric Acid Affinity Liquid Chromatography	Whole blood	5 µl	3min
ACR	mAlb	Urinary Microalbumin	Fluorescence	Urine	150 µl	8min
	Ucr	Creatinine	Benedict-Behre method			
	mAlb/Ucr	mAlb/Ucr	/			
Digestive System Screening	TBA	Total Bile Acid	Rate method	Serum	150 µl	10min
	CHE	Cholinesterase				
	α-AMY	α-Amylase				
Homocysteine	HCY	Homocysteine	Enzymatic cycling method	Serum	150 µl	9min
C-Reactive Protein / Anti-streptolysin O / Rheumatoid Factor	CRP	C-Reactive Protein	Latex immunoturbidimetry	Serum	150 µl	13min
	ASO	Anti-streptolysin O				
	RF	Rheumatoid Factor				
Immunoglobulin	IgA	Immunoglobulin A	Immunoturbidimetry	Serum	150 µl	13min
	IgG	Immunoglobulin G				
	IgM	Immunoglobulin M				
Electrolyte (I)	K	Potassium	Rate method/end-point method	Serum	150 µl	13min
	Na	Sodium				
	Cl	Chlorine				
	Ca	Calcium				
Electrolyte (II)	P	Phosphorus	Rate method/end-point method	Serum	150 µl	15min
	Mg	Magnesium				
	Fe	Iron				
	CO ₂	Carbon dioxide				